

Ship To:  
PO DI CEVA  
Via San Bernardino 4  
Ceva  
Italy

Test and Verification Report

Case Number: 0124151791  
Work Order: WO-11254251

WORK DETAILS

Engineer Name: Eric Perosino  
Device Status: Pass  
Event Type: Predictive / Preventative Maintenance  
Service Activity:

EQUIPMENT DETAILS

Product Name: Brilliance CT 16 Slice  
Product Number: 728246  
Tech ID: IT62138170  
Serial Number: 260035  
Installed Product: 62138170  
UDI:

	Min	Max	UoM	Measurement	Result	Comments
--	-----	-----	-----	-------------	--------	----------

Electrical Safety CT

Ground Leakage Current

Ground Leakage Current	5	mA	2	Pass
------------------------	---	----	---	------

Protective Earth

Gantry Frame	200	mOhm	58	Pass
Gantry Rotor	200	mOhm	61	Pass
Patient Table	200	mOhm	52	Pass
CT Host Cabinet	200	mOhm	63	Pass
Recon Cabinet	200	mOhm	63	Pass
Transformer	200	mOhm	59	Pass
UPS	200	mOhm	51	Pass
Monitor	200	mOhm	50	Pass

Philips Representative Signature:



Engineer Name: Eric Perosino  
Date: Nov 28, 2025

